

ਮਾਤਾ ਕੁੱਸ਼ਲਿਆ ਸਰਕਾਰੀ ਹਸਪਤਾਲ, ਪਟਿਆਲਾ MATA KAUSHALYA GOVERNMENT HOSPITAL, PATIALA

ਦਫਤਰ: ਡਿਪਟੀ ਡਾਇਰੈਕਟਰ -ਕਮ- ਮੈਡੀਕਲ ਸੁਪਰਡੈਂਟ
O/o. Deputy Director -cum- Medical Superintendent

ਹਵਾਲਾ ਨੰ/Ref. No: MKH/PTL/_____
ਮਿਤੀ/Date: _____



ਟੈਲੀਫੋਨ/TeleFax: +91 175 22 22 481
ਈਮੇਲ/eMail: mkhpatiala@gmail.com

Sub: Tender Purchase for Hospital Cylindrical Autoclave (Horizontal) at MKH Patiala

Please Quote your minimum-rates for **Hospital Cylindrical Autoclave (Horizontal)** mentioned as per below required for the Medical Superintendent, M.K.H., Patiala to reach this office on or before 04/01/2024 before 17:00 PM Tender received after the due date will not be considered. The Tender must be sealed.

Name of Item: **Hospital Cylindrical Autoclave (Horizontal)**

- **Construction:-** Material SS 304
- **Wall Type:-** Double walled
- **Design Type:-** Cylindrical
- **Working Pressure:-** 20 PSI
- **Sterilization Period:-** 20 Minute
- **Volume (Capacity):-** 750 Liter
- **Height:-** 600 millimeter
- **Diameter:-** 600 millimeter
- **Depth:-** 1200 millimeter
- **Width:-** 450 millimeter

Note – Unsealed quotations will not be considered.

Terms and Conditions:-

1. The rates must be quoted for Office of Medical Superintendent, Mata Kaushalya Hospital, Near District Court, Lahori Gate, Patiala.
2. The conditional quotations will not be accepted.
3. The Quotations will be valid for 180 days from the date of open, can be extended.
4. The quantity can be increased or decreased as per actual requirement at the time of placement of order.
5. Medical Superintendent, Mata Kaushalya Hospital, Near District Court, Lahori Gate, Patiala reserves the right to accept or reject any or all the tenders without assigning any reasons.
6. Mention the company Name of Items.